

MEMSSTAR User Instructions



**University of Notre Dame
Department of Electrical Engineering**

1. Sign into the Coral and enable the machine.
2. On the back of the machine, locate the pump control handheld unit.
 - a. Press "Enter" to clear warnings on the controller. Yellow warning light will go from flashing to steady on when all are acknowledged.
 - b. Once warnings are acknowledged, press the "Start" button on the controller.
 - c. Press "Enter" to start the pump.
3. Reset EMO pushbuttons.
4. On back of the machine, push the white "Reset" button. The green "Start" button should lite.
 - a. Push the green "Start" button to start the system.
5. Open the machine right panel labelled "Computer" and depress the computer power button to start the computer. Allow the computer to boot up and into windows.
6. As needed, press the "Login" button on the computer and enter user name(super) and password (super).
7. Press "OK" with the mouse.
8. Select the recipe "01 Pump Purge Vent" and press "Load".
9. Press "Run" to run this process.
10. Follow any on screen instructions as needed.
11. When recipe is completed and the system is "IDLE" again.
12. Open Lid and load samples into the chamber.
13. Select your recipe from the list and press "Load".
14. Press "Run" on screen when ready to start processing.
15. System will take over and follow on screen prompts if needed. Step information is listed in the lower right side of the screen.
16. When process is complete the chamber should be at atmosphere and ready to unload or change your samples.
17. Perform steps #5 - #14 for as many recipes as needed on your samples.
18. "Load" recipe "01 Chamber Pump Down"
19. "Run" this recipe.
20. After this recipe has completed and the machine is back in IDLE.
21. On the computer keyboard, depress the windows key and shut down the computer.
22. On the Handheld pump control.
 - a. Press the "Stop" button.
 - b. Press "Enter" to stop pump
 - c. Default option should be "Auto Shutdown" but select it if needed.
 - d. Press "Enter" to start automatic pump shutdown.
23. After the computer shuts down, depress and EMO switch to shut down the system.
24. Disable the machine in Coral and fill in the Run Data Collector.

Figure #1: Handheld pump Control

